

PT-06: "DETECTION OF ANTI-TOXOPLASMA IgG IN ANIMAL SERUM SAMPLES"

Instructions

The same day of items shipping, the participant receives a link to an on-line form where the following information must be reported:

- Package content and its condition of preservation
- Materials and Methods used to analyze PT samples
- Results

The on-line form remain active up to the due date (specified in the PT request form), after this date, results will not be accepted.

At arrival in the lab, the packaging and its contents must be checked for correctness and completeness.

Before performing the test, the following remarks are to be taken into account:

1. it's necessary to treat PT items in the same manner as the routinely tested samples;
2. the samples have to be stored refrigerated at +4-+15°C until the test is performed;
3. to detect anti-*Toxoplasma* IgG in serum samples, labs may use any serological test based on the detection of anti-*Toxoplasma* IgG . Each laboratory may choose the test/s routinely used;
4. any variation from what reported in the kit manual must be described in the on-line form;
5. whenever not expressly requested by the kit, each serum sample have to be tested in duplicated;
6. samples have to be handled by the personnel following the routine safety procedures requested for infectious biological material, i.e. wearing individual protection devices (coat, mask and gloves). Specific safety measures must be followed according to the test procedure applied.

For any information or problem related to the PT participation, please address to:

Dr. Alessandra Ludovisi; e-mail: alessandra.ludovisi@iss.it