

European Union Reference Laboratory for Parasites
Department of Infectious Diseases
Unit of Foodborne and Neglected Parasitic Diseases



Istituto Superiore di Sanità

Detection of anti-Trichinella antibodies in swine serum by indirect **ELISA**

INDEX

1.	Aim and field of application	2		
2.	Principle of the method	2		
3.	References	2		
4.	Acronyms	2		
5.	Devices/instruments	2		
6.	Reagents and chemicals	3		
7.	Procedure			
	7.1 Test and control samples preparation	4		
	7.2 Procedure	4		
8.	Interpretation of the results	5		
9.	Characteristics of the method	5		
10.	10. Safety measures			

ANNEX 1 - Production of excretory/secretory antigens from muscle larvae of Trichinella spiralis

rev. 9, 2024 page 1 of 8

OLVIIIRS OF SWALL

European Union Reference Laboratory for Parasites

Department of Infectious Diseases Unit of Foodborne and Neglected Parasitic Diseases





1. Aim and field of application

To determine the presence of anti-*Trichinella* antibodies by an enzyme linked immunosorbent assay (ELISA) in swine sera.

The method is used to screen pig populations during seroepidemiological surveys on the exposure of swine to *Trichinella*.

2. Principle of the method

A 96-well microtiter polystyrene plate is coated with *Trichinella spiralis* excretory/secretory antigens (E/S Ag) partially purified in conditions able of maintaining the antigens' native form.

Control and test serum samples, properly diluted, are distributed in the wells, allowing any anti-*Trichinella* sp. antibodies that are present to bind to the adsorbed antigen.

The antibodies that do not bind to the antigen are removed by washing; peroxydase conjugated antiswine IgG rabbit antibody is then added to each well. This second incubation allows the conjugate to bind to the swine antibodies that were bound to the antigens onto the well surface.

The excess conjugate is removed by washing, and the activity of the enzyme bound to the swine antibodies is measured by adding a chromogen substrate. After incubation, the intensity of the developed color is determined by a spectrophotometer.

The result is interpreted comparing the color intensity of the wells containing the test sera with those containing the controls.

3. References

Gamble HR, Anderson WR, Graham CE, Murrell KD. 1983. Diagnosis of swine Trichinellosis by enzymelinked immunosorbent assay (ELISA) using an excretory-secretory antigen. Vet. Parasitol. 13, 349-361.

OIE/World Organisation for Animal Health, 2018. Trichinellosis (infection with *Trichinella* spp.), Chapter 3.1.20. Manual of Diagnostic Tests and Vaccines for Terrestrial Animals. World Organization for Animal Health www.oie.int/eng/normes/mmanual/A_00013.htm.

Gómez-Morales MA, Ludovisi A, Pezzotti P, Amati M, Cherchi S, Lalle M, Pecoraro F, Pozio E; the Ring Trial Participants. 2009. International ring trial to detect anti-*Trichinella* IgG by ELISA on pig sera. Vet Parasitol. 166:241-248.

Bruschi F, Gómez-Morales MA, Hill D. 2019. International Commission on Trichinellosis: Recommendations on the use of serological tests for the detection of *Trichinella* infection in animals and humans. Food and Waterborne Parasitology, https://doi.org/10.1016/j.fawpar.2018.e00032.

4. Acronyms

ELISA Enzyme Linked Immunosorbent Assay

Ag Antigen
Ab Antibodies

E/S Ag Excretory/Secretory antigens
BSA Bovine Serum Albumin

5. Devices/instruments

The following instruments are needed to perform the ELISA:

Adjustable pipettes (volumes: 0.5-10 μL, 15-300 μL, 5-1000 μL)

Scale 0.001÷310g; 0.001/0.01g resolution

Automatic plate washer (strongly recommended)

Microtiter plate spectophotometer, 450 nm

Freezer ≤ -15°C

Freezer ≤ - 50°C

Refrigerator 1÷8°C

rev. 9, 2024 page 2 of 8

OLAMIAS OF SWAN

European Union Reference Laboratory for Parasites

Department of Infectious Diseases
Unit of Foodborne and Neglected Parasitic Diseases

Istituto Superiore di Sanità



Incubator 37±1°C

Magnetic stirrer

Adjustable volume dispenser, alternatively pipettes could be used

pH meter

Vortex

6. Reagents and chemicals

The step-by-step procedure for preparing the reagents is described below.

- 6.1 Analytical grade water
- 6.2 Phosphate buffered saline (PBS), pH 7.3±0.2

 KH_2PO_4 0.34 g

Na₂HPO₄ 1.21 g

NaCl 8.0 g

Analytical grade water up to 1000 mL

Dissolve compounds in 750 mL of analytical grade water under magnetic stirring.

Check the pH (7.3±0.2) and bring the solution to the final volume. Store at 1÷8°C. Stability 6 months.

6.3 Carbonate buffered saline, pH 9.6±0.2

Na₂CO₃ 1.12 g NaHCO₃ 2.92 g

Analytical grade water up to 1000 mL

Dissolve the compounds in 750 mL of analytical grade water under magnetic stirring.

Check the pH (9.6±0.2) and bring the solution to the final volume. Store at room temperature. Stability 6 months.

6.4 Washing solution

Tween 20 1 mL
Analytical grade water up to 2000 mL

Prepare the solution immediately before use, as follows: add 1 mL of Tween 20 into a 2 L flask; bring the solution to the final volume by adding analytical grade water and mix by magnetic stirring until the solution is clear. If refrigerated (1÷8°C), stability 24 h.

6.5 Blocking solution

BSA 0.25 g Tween 20 0.025 mL PBS up to 50.00 mL

Prepare the solution immediately before use, as follows: place 0.25 g BSA (bovine serum albumin) directly in a 50 mL tube; add 40 ml of PBS buffer and mix by vortexing until the BSA is completely dissolved. Add 0.025 mL Tween 20; mix by vortexing and bring to volume. If refrigerated (1÷8°C), stability 24 h.

6.6 Sera and conjugate diluter

BSA 0.50 g Tween 20 0.025 mL PBS up to 50 mL

Prepare the solution immediately before use, as follows: place 0.50 g BSA directly in a 50 mL tube; add 40 ml of PBS buffer and mix by vortexing until BSA is completely dissolved. Add 0.025 mL Tween 20; stir by vortexing and bring it to volume. If refrigerated (1÷8°C), stability 24 h.

6.7 Stop solution

HCl 1N in analytical grade water. Prepare the solution under a chemical hood; store at room temperature. Stability 6 months.

6.8 TMB (3, 3', 5, 5' tetramethylbenzidine) peroxidase substrate

rev. 9, 2024 page 3 of 8

SALEMOK DI SAMILA

European Union Reference Laboratory for Parasites

Department of Infectious Diseases
Unit of Foodborne and Neglected Parasitic Diseases

Istituto Superiore di Sanità



This substrate is recommended; if not available, any other peroxidase substrate can be used.

6.9 96-well flat-bottomed microtiter plate

6.10 Excretory/secretory antigens (E/S Ag) (see ANNEX 1)

Reconstitute the antigens using analytical grade water to 25-50 μ g/vial. Reconstituted E/S Ag, frozen at <-50°C is stable for 10 years. Dilute the reconstituted antigens with carbonate buffer saline pH 9.6, to a final concentration of 5 μ g/mL just before use.

6.11 Peroxidase labelled anti-swine IgG goat antibodies

On opening the packaging, rehydrate the freeze-dried material with analitycal grade water shaking on vortex until its complete dissolution.

The solution can be stored refrigerated (1÷8°C) for a week.

To determine the optimal working dilution of the conjugate, a titration curve is performed by testing working dilutions higher, lower and equal than those recommended by the supplier for use in ELISA. The working dilution of the conjugate is the dilution in which the differences in optical density, O.D., between positive and negative controls are maximum maintaining the minimum background, set by the O.D. value of the blank.

The conjugate, once rehydrate, if distributed in aliquots and stored frozen at <-50°C remains stable for at least 20 years. After the expiration date, its suitability is verified, through the O.D. values detected in the positive and negative controls, in the analytical sessions in which it is used.

To perform the test, dilute the conjugate aliquot to the optimum concentration with the dilution solution referred to in point 6.6. Once diluted, store the conjugate refrigerated (1÷8°C). Stability 24 hours.

6.12 Anti-Trichinella sp. positive control sera

Serum coming from pig sperimentally infected with *Trichinella* sp. larvae (positive control). Positive sera are considered as reference material of the European Union Reference Laboratory for Parasites (EURLP).

6.13 Anti-Trichinella sp. negative control sera

Serum coming from domestic pig tested negative for the presence of *Trichinella* sp. by peptic chloride.

7. Procedure

7.1 Test and control samples preparation

Thaw the test sera and the positive and negative control sera by storing them at 1÷8°C for at least 5 h.

Once thawed, keep them in an ice bath and stir them by vortexing before use.

Dilute 1:50 the test and control sera as follows: in a 1-2 mL conical bottom tube add 10 µL of serum and 490 µL diluting solution. Diluted sera can be stored refrigerated (1÷8°C), stability 24 h.

7.2. Procedure

Fill, each well of the microtiter plate, with 100 μL of E/S Ag in carbonate buffered saline; incubate for 1h at 37±1°C.

Wash 3 times in the automatic plate washer with the washing solution.

Dispense, in each well, 200 μ L blocking solution per well; incubate for 1 h at 37° C.

Wash 3 times in the automatic plate washer with the washing solution.

Dispense, in duplicate, 100µL per well of each serum (negative, positive controls and testing samples).

Dispense 100µL of sera and conjugate diluter solution (6.6) in the blank wells.

Incubate per 30 minutes at 37±1°C

Wash 3 times in the automatic plate washer with the washing solution.

Add $100\mu L$ to each well of the diluted anti-swine IgG peroxidase labelled antibodies and incubate for 1h at $37\pm1^{\circ}C$.

Wash 3 times in the automatic plate washer with the washing solution.

Add 100µL TMB substrate to each well; incubate for 10 min at room temperature.

rev. 9, 2024 page 4 of 8



European Union Reference Laboratory for Parasites

Department of Infectious Diseases
Unit of Foodborne and Neglected Parasitic Diseases



Istituto Superiore di Sanità

Stop the reaction by adding 50μ L of the stop solution to each well and read the reaction in the microtiter plate spectophotometer at 450nm.

8. Interpretation of the results

The test results can be considered as valid if all of the following criteria are fulfilled:

The O.D. value of the negative control sera is lower than the *cut off* value determined by the validation method.

The OD value of the positive control sera has to be >1.0 of the unit absorbance.

The difference, in O.D., between the two measures made on the same positive control serum in strict conditions of repeatability is \leq 0.15 unit of absorbance;

The difference, in O.D., between the two measures made on the same negative control serum is ≤ 0.05 unit absorbance.

If even only one of the above-reported criteria is not met, the test has to be considered as non-valid and the sera should be tested again.

To evaluate the obtained results:

Calculate the mean of the two duplicates for each positive sera (PS) and for each test sera (TS).

Subtract from each mean value the mean OD value of the blanks (OD_b).

Select the higher OD value among the positive control sera (PS_{max}), and for each sample calculate the extinction value (I_e) according to the following formula:

	OD mean duplicates TS – OD _b
I _e (%) =	X 100%
	OD mean duplicates highest PS – OD _b

Test result is POSITIVE (presence of anti- *Trichinella* antibodies) if the extinction value $I_e \ge 18\%$.

Test result is NEGATIVE (absence of anti- *Trichinella* antibodies) if the extinction value I_e < 18%.

9. Characteristics of the method

This method was validated in terms of diagnostic sensitivity, specificity and repeatability. The results of the validation procedure confirmed that the method is suitable for the expected aim and field of application.

10. Safety measures

This method has to be carried out only by authorized personnel. The operator should wear individual protection devices during the test performance. For the general safety measures, refer to the guidelines of CDC.

rev. 9, 2024 page 5 of 8



European Union Reference Laboratory for Parasites

Department of Infectious Diseases
Unit of Foodborne and Neglected Parasitic Diseases

Istituto Superiore di Sanità



ANNEX 1

Production of excretory/secretory antigens from muscle larvae of Trichinella spiralis

1 Aim and field of application

To describe the production of excretory/secretory antigens from muscle larvae of Trichinella spiralis.

The product can be used as antigens for serology to detect anti-Trichinella spp. antibodies.

2 References

Gamble HR, Anderson WR, Grahan CE, Murrell KD. 1983. Diagnosis of swine trichinellosis by enzyme-linked immunosorbent assay (ELISA) using an excretory-secretory antigen. Vet. Parasitol. 13, 349-361

Gamble HR, Rapic D, Marinculic A, Murrell KD. 1988. Evaluation of excretory-secretory antigens for the serodiagnosis of swine trichinellosis. Vet. Parasitol. 30, 131-137

OIE/World Organisation for Animal Health, 2018. Trichinellosis (infection with *Trichinella* spp.), Chapter 3.1.20. Manual of Diagnostic Tests and Vaccines for Terrestrial Animals. World Organization for Animal Health www.oie.int/eng/normes/mmanual/A_00013.htm.

3 Acronyms

OD, optical density

E/S Ag, Excretory/Secretory antigens

4 Devices/Instruments

Incubator 37 ± 2°C with 4-5% CO2

Spectrophotometer UV/VIS

Freezer ≤-15°C

Refrigerator, 1°C ÷ 8°C

Freezer < -50°C

Laminar flow hood

Adjustable micropipettes (up to 2 μ L, 20 μ L, 200 μ L, 1000 μ L)

Pipettes (1, 5, 10, 25 mL)

Inverted microscope

Magnetic stirrer

Refrigerate centrifuge

96 wells plates

5. Reagents and chemicals

5.1 Phosphate Buffered Saline (PBS), pH 7.3±0.2

 $\begin{array}{ccc} \text{KH}_2\text{PO}_4 & 0.34 \text{ g} \\ \text{Na}_2\text{HPO}_4 & 1.21 \text{ g} \\ \text{NaCl} & 8.0 \text{ g} \\ \text{Analytical grade water} & \text{up to 1000 mL} \end{array}$

Dissolve the compounds in 750 mL of analytical grade water under magnetic stirring. Control the pH (it must be 7.3 \pm 0.2) and then bring the solution to the final volume. Sterilize by filtration with a 0.22 μ m filter. If stored refrigerated, the solution must be used within 6 months.

5.2 PBS, pH 7.3±0.2 with antibiotics 5 X

PBS 950 mL

Penicillin/Streptomycin or Antibiotic/Antimycotic solution

50 mL

Store refrigerated and use within 2 months.

rev. 9, 2024 page 6 of 8

OLALIUS OLALIUS

European Union Reference Laboratory for Parasites

Department of Infectious Diseases
Unit of Foodborne and Neglected Parasitic Diseases





5.3	RPMI	1640	with	antibiotics	5 X
J.J	1 1 1 1 1 1 1 1	1040	WILLI	antibiotics	σ

RPMI 1640	475 mL
Penicillin/Streptomycin or Antibiotic/Antimycotic solution	25 mL

Store refrigerated and use within 2 months.

5.4 Complete RPMI medium

RPMI 1640	480 mL
1M HEPES	5 mL
200mM Glutamine	5 mL
100mM Na-pyruvate	5 mL
Penicillin/Streptomycin or Antibiotic/Antimycotic	

Store refrigerated and use within 2 months.

- 5.5 Penicillin/Streptomycin or Antibiotic/Antimycotic solution (100x).
- 5.6 RPMI 1640 culture medium.
- 5.7 HEPES-1M Buffer solution.
- 5.8 L-Glutamine 0.2M solution.
- 5.9 Sterile 0.22 μm syringe filters.
- 5.10 Device for concentrating by ultrafiltration with cellulose membrane filter, 5000 MWCO cut off (e.g. Amicon Ultra 15).
- 5.11 Coomassie Protein Assay Reagent (e.g. Pierce ®, 1856209).
- 5.12 Slide-A-Lyzer Dialysis Cassettes, 3.5K MWCO 3-12 ml Capacity (e.g. Pierce ®, 66110).
- 5.13 Cocktail of protease inhibitors.
- 5.14 Trichinella spiralis muscle larvae suspension (MLS).
- 5.15 Sterile conical tubes 15, 50 mL.
- 5.16 Tissue culture flasks T-75.

6 Procedure

- a) Place the solutions 5.1, 5.2 and 5.3 into the 37°C incubator for at least 1 h.
- b) Count MSL (5.14) under the inverted microscope.
- c) Wash 5x10⁵ MSL (5.14) three times (20 min each wash) by sedimentation in a sterile 50 ml conical tube with 45 mL of warm sterile PBS/Penicillin/Streptomycin solution (5.2). At each change of the washing solution, gently shake the worms to dislodge adherent bacteria. After larvae sedimentation, remove the washing solution by aspirating with a pipette.
- d) Place the tube containing the worms (5.14) under a laminar flow hood and wash them 5 additional times by sedimentation in a sterile 50 mL conical tube with 45 mL of warm RPMI 1640/Penicillin/Streptomycin solution (5.2).
- e) Resuspend the worms (5.14) in warm maintenance media (5.4) at a concentration of 5,000 larvae/mL and place them into T-75 tissue culture flasks.
- f) Incubate the flasks in 5% CO₂ at 37°C for 16-18 h.
- g) Check MSL (5.14) viability and the absence of bacterial contamination by microscopy.
- h) Place the cultures under a laminar flow hood, let the MSL sediment and transfer the culture media in 50 mL conical tubes.
- i) Filter the media (E/S Ag) through a 0.22 μm filter (5.9). Discard the MSL.
- j) Keep the E/S Ag at +4°C until concentration, if stored more than 24 hours, the E/S Ag should be frozen at -15°C.
- k) Fill in the concentration tube (5.10) with 15 mL of E/S Ag (point i).
- I) Centrifuge 30 minutes at 3,000 g in a refrigerated centrifuge.
- m) Retrieve the concentrated E/S Ag in a 50 mL tube and store refrigerated until point "p".

rev. 9, 2024 page 7 of 8

OLALING OLANIA

European Union Reference Laboratory for Parasites

Department of Infectious Diseases
Unit of Foodborne and Neglected Parasitic Diseases

Foodborne and Neglected Parasitic Diseases Istituto Superiore di Sanità



- n) Repeat point "k" to "m" until the E/S Ag has been properly concentrated.
- o) If the E/S Ag has been concentrated 100 times or more, proceed to point "p", otherwise centrifuge the tube again.
- p) Determine the protein concentration by the Bradford method (5.11).
- q) Check the O.D. by the spectrophotometer at 280nm/260nm. The ratio should be ≥ 1 .
- r) Dialyze (5.12) the concentrated E/S Ag versus PBS at +4°C for at least 4 hours.
- s) Add to the E/S Ag 1µL/mg of the cocktail of protease inhibitors (5.13).
- t) Aliquot the antigen in 1 mL vials at a total protein concentration of 0.5 mg/vial and store at ≤-15°C.
- u) The antigen can be lyophilized and stored refrigerated for 5 years.

7. Quality Control

The batch is considered suitable for serology if all the following criteria are fulfilled:

Lack of microbial and fungal contamination in the culture flasks at the end of the incubation period, verified by direct microscopic observation at 400x magnifications, as specified in point 6 'g';

MSL viability, controlled as described in point 6 'g';

O.D. ratio of 280nm/260nm of the final antigen solution ≥ 1, as specified in point 6 'q.'

8. Safety

This method has to be carried out only by authorized personnel. The operator should wear individual protection devices during the test performance. For the general safety measures, refer to the guidelines of CDC.

rev. 9, 2024 page 8 of 8