

European Union Reference Laboratory for Parasites
Istituto Superiore di Sanità

Proficiency Testing for the detection of anti-*Toxoplasma* IgG in ovine serum samples

Seventeenth Annual EURLP Workshop, Rome 15-16 September 2022

Alessandra Ludovisi



AIMS

The aims of the proficiency testing (PT) were :

To evaluate:


- the laboratories competence in detection of anti-*Toxoplasma* IgG in ovine serum samples
- the performance of the commercial kits, utilized for the detection of anti-*Toxoplasma* IgG in ovine.

Participating laboratories

Austria 

Belgium 

Denmark 

Germany (2) 

Greece 

Iceland 

Italy (2) 

Lithuania 

Latvia 

Netherland 

Norway 

Poland 

Portugal 

Romania 

Slovak 

Slovenia 

Spain 

Methods

Any serological test based on the detection of anti-*Toxoplasma* IgG could be used. Each laboratory should choose the test/s routinely used.

Materials and Methods 1.

The test material forwarded to each laboratory consisted of a panel of 4 serum samples collected from *T. gondii* free or *T.gondii* experimentally infected animals.

Serum sample panel

Serum sample Code	Expected Result	S/P%*
s.1	Positive	> 100%
s.2	Positive	82.8%
s.3	Positive	59.1%
s.4	Negative	4.2%

*These serum samples were:

individually tested for anti-*Toxoplasma* IgG by the: “ID Screen Toxoplasmosis Indirect Multi-species

Materials and Methods 2.

Serum samples were:

- **distributed in 100 μ L aliquots,**
- **preserved with 1% merthiolate solution,**
- **labelled with a code**

Materials and Methods 3.

Each laboratory received a link available to NRLs on the EURLP website to the online survey platform Microsoft Forms, through which PT results and other details could be submitted



Results1.

The methods or kits used by the participants were:

- **IDVet ID Screen, 15 labs:** (A,C,D,E,G,H,L,M,N,O,P,Q,R,S,T)
- **Indirect Immunofluorescence antibody test (IFAT) / OIE, 1 lab:** F
- **Toxoreagent Mast Kit, 2 labs:** B and I
- **VetLine Toxoplasma ELISA (NovaTec, Immundiagnostica GmbH), 1 lab:** J

Two labs performed an additional test:

- **Immunoblot full Antigen and P30 Antigen, 1 lab :** E
- **Pigtype Toxoplasma Ab, Indical Bioscience, 1 lab :** A

Criteria for the results evaluation:

The participating laboratory had to indicate the positivity or negativity of each tested item.

The result of the analysis of each serum sample was reported as



correctly classified



incorrectly classified

RESULTS 2.

11/19 NRLs correctly classified (as positive or negative)
all serum samples

Sample code	Expec. results	Lab. code																		
		A	E	F	H	I	M	N	O	P	Q	S	B	C	D	G	J	L	R	T
1	+	+	+	+	+	+	+	+	+	+	+	+	-	+	+	+	+	+	+	+
2	+	+	+	+	+	+	+	+	+	+	+	+	-	+	+	+/-	-	+	+	+
3	+	+	+	+	+	+	+	+	+	+	+	+	-	+/-	+/-	-	-	+/-	-	+/-
4	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-

Test interpretation

S/P \leq 40% Negative
40% < S/P < 50% Doubtful
S/P \geq 50% Positive

—————> IDvet



Serum sample Code	Expected Result	S/P
s.1	positive	152%
s.2	positive	82.8%
s.3	positive	59.1%
s.4	negative	4.2%

Taking into consideration:

- Serum sample n°3 is clearly positive sample
- Aliquots were homogeneous because they were prepared from the same batch

Possible causes of disagreement:



- **Not uniform distribution of the antigen in the plate**
- **Not appropriate resuspension of all Kit components before use**
- **Human error**

Criteria for the final PT evaluation

The result evaluation is qualitative,

IgG titers will be considered as additional information to compare the performance of tests used by participants

Final evaluation is considered as

“positive” if all samples are correctly classified

“negative” in all the other cases.

PT evaluation


Laboratory code	N° of samples correctly classified	N° of samples NOT correctly classified	Final evaluation
A	4/4	0/4	Positive
E	4/4	0/4	Positive
F	4/4	0/4	Positive
H	4/4	0/4	Positive
I	4/4	0/4	Positive
M	4/4	0/4	Positive
N	4/4	0/4	Positive
O	4/4	0/4	Positive
P	4/4	0/4	Positive
Q	4/4	0/4	Positive
S	4/4	0/4	Positive
B	1/4	3/4	Negative
C	3/4	1/4	Negative
D	3/4	1/4	Negative
G	2/4	2/4	Negative
J	2/4	2/4	Negative
L	3/4	1/4	Negative
R	3/4	1/4	Negative
T	3/4	1/4	Negative

Conclusion

More than a half (58%) of the laboratories correctly classified all serum samples

The more frequently used commercial kits are immunoenzymatic assays

Overtime comparison





Thanks for your attention!