

***Towards novel validated PCR methods
and innovative reference materials
for the detection of VTEC *Escherichia coli****

***5th Annual workshop of the National Reference Laboratories for E. coli
Rome, 08 October 2010***




How to validate an analytical method?

By performing an in-house validation
(pre-validation)


By conducting a collaborative study (full-validation)

ISO 16654:2001 for the detection of *E. coli* O157 in food applying immuno-magnetic separation consisting of several time-consuming steps



“Technical Specification” by Working Group 6 of the Technical Committee 275 of the European Normalisation Committee (CEN TC275/WG6); PCR-based

Day 0 ➤ Day 1



Method Modularity

A model that allows flexibility and enforcement

- We applied the concept called “Method Modularity”
- The PCR “module” is applicable to any DNA template containing a given target.
Provided that the template DNA is prepared by suitable methods, and fully characterised in terms of quality/quantity prior to use in PCR.

PCR methods

Target	Forward primer, reverse primer and probe sequences (5'-3')	Amplicon size (bp)	Location within sequence	GenBank accession number
stx1 (Perelle S. et al. Mol Cell Probes 2004 18:185–192)	TTTGTYACTGTSACAGCWGAAGCYTTACG CCCCAGTTCARWGTRAGRTCMACRTC Probe -CTGGATGATCTCAGTGGCGTTCTATGTAA	131	878–906 983–1008 941–971	M16625
stx2 (Perelle S. et al. Mol Cell Probes 2004 18:185–192)	TTTGTYACTGTSACAGCWGAAGCYTTACG CCCCAGTTCARWGTRAGRTCMACRTC Probe -TCGTCAGGCAGTGTCTGAAACTGCTCC	128	785–813 785–813 838–864	X07865
eae (Møller Nielsen E. and Thorup Andersen M. J clin Microbiol 2003)	CAT TGA TCA GGA TTT TTC TGG TGA TA CTC ATG CGG AAA TAG CCG TTA Probe -ATAGTCTGCCAGTATTGCCACCAATACC	102	899–924 1000–979 966–936	Z11541
rfbE (O157) (Perelle S. et al. Mol Cell Probes 2004 18:185–192)	TTTCACACTTATTGGATGGTCTCAA CGATGAGTTTATCTGCAAGGTGAT Probe -AGGACCGCAGAGGAAAGAGAGGAATTAAGG	88	348–372 412–435 381–410	AF163329
wbdI (O111) (Perelle S. et al. Mol Cell Probes 2004 18:185–192)	CGAGGCAACACATTATATAGTGCTTT TTTTGAATAGTTATGAACATCTTGTAGC Probe -TTGAATCTCCCAGATGATCACACATCGTGAA	146	3464–3489 3579–3609 3519–3548	AF078736
§ wzx (O26) (Perelle S. et al. Mol Cell Probes 2004 18:185–192)	CGCGACGGCAGAGAAAATT AGCAGGCTTTATATTCTCCAACCTT Probe -CCCCGTTAACATCAATACTATTCACGAGGTTGA	135	5648–5666 5757–5782 5692–5724	AF529080
ihp1 (O145) (Perelle S. et al. Mol Cell Probes 2004 18:185–192)	CGATAATATTACCCCACCACTACAG GCCGCCGCAATGCTT Probe -CCGCCATTAGAATGCACACAATATCG	132	1383–1408 1500–1514 1472–1498	AF531429
wzx (O103) (Perelle S. et al. J Appl Microbiol 2005 98:1162–1168)	CAAGGTGATTACGAAAATGCATGT GAAAAAAAGCACCCCCGTACTTAT Probe -CATAGCCTGTTGTTTAT	99	4299–4323 4397–4375 4356–4373	AY532664

Plasmid collection

pCRL-ECstx1a

5' TTTGTTACTGTGACAGCTGAAGCTTACGTTTCGGCAAATACAGAGG
 GGATTTCGTACAAACACTGGATGATCTCAGTGGCGTTCTTATGTAATGA
 CTGCTGAAGATGTTGATCTTACATTGAACCTGGGG 3'

pCRL-ECstx2a

5' TTTGTCACTGTCACAGCAGAACGCCTACGCTTCAGGCAGATAACAGA
 GAGAATTTCGTCAGGCAGTGTGAAACTGCTCCCTGTGTATACGATGA
 CGCCGGGAGACGTGGACCTCACTCTGAACCTGGGG 3'

pCRL-ECstx2b

5' TTTGTCACTGTCACAGCAGAACGCCTACGGTTTCAGGCAGATAACAGA
 AGAATTTCGTCAGGCAGTGTGAAACTGCTCCCTGTGTATACGATGA
 CGCGGAGAAAGTGGACCTCACACTCTGAAC 3'

pCRL-ECstx2c

5' TTTGTCACTGTCACAGCAGAACGCCTACGGTTTCAGGCAGATAACAGA
 GAGAATTTCGTCAGGCAGTGTGAAACTGCTCCCTGTGTATACGATGA
 CGCCGGGAGACGTGGACCTCACTCTGAAC 3'

pCRL-ECstx2d

5' TTTGTCACTGTCACAGCAGAACGCCTACGGTTTCAGGCAGATAACAGA
 GAGAATTTCGTCAGGCAGTGTGAAACTGCTCCCTGTGTATACGATGA
 CGCCGGGAGACGTGGACCTCACTCTGAAC 3'

pCRL-ECeae

5' CATTGATCAGGATTTCTGGTGATAATACCGTTAGGTATTGGTG
 GCGAATACTGGCGAGACTATTCAAAAGTAGCGTTAACGGCTATTCC
GCATGAG 3'

pCRL-ECrbE (O157)

5' TTTCACACTTATTGGATGGCTCAATTCTAATCAGGACCCGAGAGGA
 AAGAGAGGAATTAAAGGAATCACCTTGAGATAAACTCATCG 3'

pCRL-ECwbdI (O111)

5' CGAGGCAACACATTATATAGTGCTTTGTTACACACTGAAAGTTCTTA
 AAAGTGAATTGAATCTCCCAGATGATCAACATCGTGAATACCTTTGGCT
 AACTAAACACCAAATAAAGCTAAACAAGATGTTCATAACTATTCAAA
A 3'

pCRL-ECswzx (O26)

5' CGCGACGGCAGAGAAAATTATTAAATGTATTCACTAGTCTATAGCAACCC
 CGTTAAATCAATACTATTACGAGGTTGATAAAAGCAACATGAATTGAA
 ATTAGAACCATACAAAGTTGGAGAATATAAAAGCCTGCT 3'


pCRL-ECihp1 (O145)

5' CGATAAATATTACCCACCAAGTACAGCCGTACAGACTGACAGCACT
 GTCACCGATAAAACCAGGAGATCGCGACGCCGCAAGATAACCGC
 CATTCAAGATGCACACAATATCGCAAGCATTGCGCGGC 3'

pCRL-ECwzx (O103)

5' CAAGGTGATTACGAAAATGCATGTTTTATTGGCGTGCATCAATTAA
 ATTATCCTTCAAGCCTGTTGTTTATTATAAGTACGGGGTGCTTTTT
C 3'

Validation process



Acceptance criteria


(pre-validation requirements:

dynamic range, efficiency, specificity, LOD)

YOU ARE HERE

Performance requirements

Validation scheme




Method	Dynamic Range (gDNA copies) (3 replicates, 3 repetitions for each level)	Efficiency	Annealing temperature
stx1	1-10.000	92%	60°C
stx2	1-10.000	89%	60°C
eae	1-10.000	90%	60°C
O26	1-10.000	99%	60°C
O103	1-10.000	80%	No signal above 55°C due to probe low a.t.
O111	1-10.000	92%	60°C
O145	1-10.000	84%	60°C
O157	1-10.000	91%	60°C

Efficiency formula: $((10^{(-1/\text{slope})}) - 1) \times 100$

3,1 < Slope < 3,6

90% < Efficiency < 110%



LOD determination

(60 replicates, 2 repetitions for each level)

Method	Copy number per reaction	Average Ct±SD	Positive (%)
stx1	15	34±0.9	60/60 (100%)
	10	36.5±1.22	57/60 (95%)
	5	37.5±1.29	52/60 (87%)
	0.1	42.3±1.47	3/60 (5%)
stx2	15	34.1±0.78	60/60 (100%)
	10	36.4±1.34	59/60 (98%)
	5	38.3±1.61	56/60 (93%)
	0.1	41.4±1.07	7/60 (12%)
eae	15	34.8±1.49	60/60 (100%)
	10	35.9±1.04	60/60 (100%)
	5	38.0±0.85	59/60 (98%)
	0.1	41.1±1.94	5/60 (8%)

Method	Copy number per reaction	Average Ct±SD	Positive (%)
serogroup O26	20	36.2±0.61	60/60 (100%)
	15	38.7±0.57	59/60 (98%)
	10	37.7±1.22	58/60 (97%)
	5	40.2±1.28	58/60 (97%)
	0.1	41.5±0.08	3/60 (5%)
serogroup O103	15	36.8±0.53	58/60 (97%)
	10	38.5±0.94	60/60 (100%)
	5	38.9±0.81	59/60 (98%)
	0.1	Undetermined	0/60 (0%)
serogroup O111	20	35.11±0.61	60/60 (100%)
	15	38.4±0.98	60/60 (100%)
	10	37.9±0.97	56/60 (93%)
	5	39.7±1.07	58/60 (97%)
	0.1	41.6	1/60 (1.7%)
serogroup O145	20	35.7±0.49	60/60 (100%)
	15	37.5±1.10	60/60 (100%)
	10	36.1±0.95	59/60 (98%)
	5	39.0±1.01	57/60 (95%)
	0.1	39.9±0.56	4/60 (6.7%)
serogroup O157	20	36.5±1.14	57/60 (95%)
	15	38.6±1.01	57/60 (95%)
	10	36.4±1.09	59/60 (98%)
	5	40.1±1.58	49/60 (82%)
	0.1	Undetermined	0/60 (0%)

Summary

Method	Reaction Efficiency	Dynamic Range (gDNA copies)	LOD (gDNA copies)
stx1	92%	1-10000	10
stx2	89%	1-10000	10
eae	90%	1-10000	5
O26	99%	1-10000	5
O103	80%	1-10000	5
O111	92%	1-10000	5
O145	84%	1-10000	5
O157	91%	1-10000	10

List of strains used for specificity testing

- *Shigella boydii*
- *Shigella dysenteriae*
- *Klebsiella pneumoniae*
- *Citrobacter freundii*
- *Yersinia enterocolitica*
- *Proteus mirabilis*
- *Campylobacter jejuni*
- *Staphylococcus aureus*
- *Hafnia alvei*
- *Enterobacter sakazakii*
- *Listeria monocytogenes*
- *Salmonella Typhimurium*
- *Salmonella Seftenberg*
- *Salmonella Hadar*
- *Salmonella Enteritidis*
- *Salmonella Cerro*

Results

- Synthetic plasmid collection has been delivered and used as positive control (11 plasmids: pCRL-ECstx1a...)
- All reactions were optimized to perform at 60°C except O103 serogroup (whose probe annealed at 55°C)
- All concentration levels used to build up each Dynamic Range were found to be significantly different among them (ANOVA, Scheffé Post-hoc test)
- LOD was fixed at 10 copies (59/60 positive replicates)

Results

- Reaction Efficiency between 80% and 99%
- Dynamic Range between 1 and 10.000 copies
- Plasmids have been used as positive controls
- LOD has been calculated
- Specificity has been tested

Monica Ermolli & Dafni Maria Kagkli

European Commission – Joint Research Centre
Ispra (VA), Italy
Tel. +390332785149
Fax +390332785904
monica.ermolli@jrc.ec.europa.eu
dafni.kagkli@jrc.ec.europa.eu

