

5th Annual workshop of the EU Reference Laboratories for E. coli Rome, 8 October 2010



4th and 5th inter-laboratory study on the detection and typing of VTEC - 2010

Gaia Scavia, Clarissa Ferreri and Stefano Morabito



5th PT: VTEC Typing

Milk Samples (25 ml) spiked with

Sample 1 40 cfu/ml VTEC O103 eae+; vt2+ 10² cfu/ml *E. faecium* ATCCL565 10² cfu/ml *E. coli* ATCC35218

Sample 2 10² cfu/ml *E. faecium* ATCCL565 10² cfu/ml *E. coli* ATCC35218

Uncertainty of Measurement (ISO TS 19036:2006)

VTEC O103 *eae+; vt2+* **0,24** log cfu/ml *E. faecium* ATCCL565 **0,38** log cfu/ml *E. coli* ATCC35218 **0,22** log cfu/ml

Stability assessment (17043:2010):

Samples prepared on April the 27th and stored @ 4°C Samples assayed on April, the 27th; 29th; May the 3rd; 5th; 10th; 12th; 14th.

Homogeneity assessment (17043:2010):

Samples prepared for shipment on May the 14th and stored @ 4°C. Three sets of samples randomly selected immediatly after the preparation and assayed on the same day.

Samples shipped in refrigerated boxes on May the 17th.

All Labs received the parcels between 18 and 19 of May.

24 Labs received the samples at temperatures compliant with the ISO 7218:2007 (range 2-8°C).

- 4 Labs received the samples at temperatures >8°C.
- 4 Labs received the samples at temperatures <2°C.
- 2 Labs did not report temperature upon arrival.

5th PT: VTEC Typing

Isolated VTEC strains seeded in soft-agar borosilicate glass vials

Sample 1 VTEC O121 vt2+, eae+ Sample 2 VTEC O91 vt1+

Sample 3 VTEC O113 vt1+, vt2+ Sample 4 VTEC O145 vt1+, eae+ Sample 5 O111 eae

Sample codes have been randomly assigned to each laboratory

5th PT: VTEC Typing

Sample preparation:

Reference stocks stored in Microbanks under nitrogen vapours.

Selected samples revitalized on nutrient Agar on May the 3rd

Single colonies picked and streaked in replicates onto Nutrient Agar on

May the 4th.

One replica used for strain checks on May the 5th.

Homogeneity assessment (17043:2010):

- Samples prepared from the other replica on May the 6th and stored @ RT until shipment.
- Three sets of samples randomly selected on May the 10th and assayed in the following two days.

5th PT: VTEC Typing

Samples shipped in the same parcel of the 4th PT samples on May the 17th.

Stability assessment not required.

Some vials of Samples 1 and 2 cross-contaminated.

Rate of contamination below 1 out of three (not detected by the Homogeneity assessment).


Contamination Occurred during the preparation of vials (different combination of reportings).

Upload of PTs results via web



Partecipants to PT4 and/or PT5

- Total participating labs to PT4 and/or PT5: 32
- •E. coli NRLs: 29/32 appointed NRLs (NRL Finland participated with two labs)
- **EU Member States**: 24 / 27 (Luxembourg participated via Belgium)
- Non-EU countries : Norway, Switzerland




Evaluation of the laboratory analytical perfomances

Agreement of results with the true values (gold standard):

K test (95% C.I.)

- Kappa test. Cohen's kappa is a measure of association between two measurements of the same item.
- K test estimates the level of agreement beyond chance
- How to interpret K value:
 Fleiss J.L. (Statistical methods for rates and proportions, 1981)

Sensitivity and Specificity




Results: participation to 4th & 5th PT

PT 4th

- Total participating labs to PT4: 29
- 26 of 32 E.coli NRLs: 81%
- 22 of 27 EU Member States: 81%
- 2 countries outside EU

4th PT: Results – Gold standard value and lab reports - 1




Sample 1: O103, vtx1-, vtx2+, eae+

25 of the 29 (86%) participating labs correctly identified the results

- 2 labs failed the identification of the **O103 serogroup** specific gene in PCR and couldn't therefore isolate the strain
- 1 lab reported a **false positive vtx1** detection in PCR and genotyped as vtx1+ the isolated strain as well
- 1 lab failed to **isolate the strain** after correct detection by PCR of the virulence and serogroup specific antigens

4th PT: Results – Gold standard value and lab reports - 2




Sample 2: negative result (PCR: vtx1-; vtx2-; eae-)

27 of the 29 participating labs (93%) correctly identified the sample


- 2 labs reported the **isolation of a O103 strain, vtx1+, vtx2-, eae+** and reported the false positive detection in PCR of O103 serogroup specific gene, vtx1 and eae genes
- possibly due to cross contamination during the analysis.

4th PT : Overall performances - 1



Of the 29 participating labs:

- 24 (83%) correctly identified both the samples
- 4 (14%) reported errors for one of the two samples
- 1 reported errors for both the samples



Of the 27 EU Member States:

• 19 (70%) had at least one NRL able to participate in the PT 4 and correctly characterize the samples

4th PT : Laboratory performances - 2


	kappa	se	sp
	1	100%	100%
	1	100%	100%
	1	100%	100%
-	1	100%	100%
	1	100%	100%
	1	100%	100%
	1	100%	100%
	0,75	67%	100%
	1	100%	100%
	1	100%	100%
	1	100%	100%
	1	100%	100%
	1	100%	100%
	1	100%	100%
•	0,14	100%	73%
-	0,2	100%	64%
	1	100%	100%
	1	100%	100%
	1	100%	100%
	1	100%	100%
-	1	100%	100%
	1	100%	100%
	1	100%	100%
-	1	100%	100%
	0,75	67%	100%
	1	100%	100%
	1	100%	100%
	1	100%	100%
•	1	100%	100%

Performances of the Real Time PCR assay


Overall Kappa: 0.936

• Sensitivity: 97,7% (95% CI: 94,5% - 100%)


• Specificity: 97,8% (95% CI: 96,2% - 99,4%)



5th PT Results – Serotyping capability of the participating



5th PT Results – Gold standard value and lab reports - 1




- Sample 1: **O121**, vtx1-, vtx2+, eae+
- Sample 2: **O91**, vtx1+, vtx2-, eae-
- Sample 3: **O113**, vtx1+, vtx2+, eae-
- Sample 4: **O145**, vtx1+, vtx2-, eae+
- Sample 5: **O111**, vtx1-, vtx2-, eae+

Assessed cross contamination of samples 1 & 2



For samples 1 & 2 false results weren't taken into account !!

5th PT Results – Gold standard value and lab reports - 2



Considering only samples 3,4,5

- Correct results 88 (91%)
- False results: 8 (8,3%)

Sample 3:

- 3 labs failed to detect serogroup O113
- 2 labs reported a false negative for vtx1
- 1 lab reported a false positive for eae


Sample 4:

•1 lab failed to detect serogroup O145

Sample 5:


1 lab reported a false positive vtx1

5th PT: Overall performances (only samples 3, 4, 5)



Of the 32 participating labs:

- 25 (76%) correctly identified all the samples
- 7 (21%) reported false results for one samples




Of the 27 EU **Member States**:

• 19 (70%) had at least one NRL able to participate in the PT 5 and correctly typed all the strains

5th PT: Lab performances (only samples 3, 4,5)

Agreement – Kappa value

Overall Kappa: 0,98 (95% CI 0,92 - 1,0)






5th PT: Lab performances (only samples 3, 4, 5)


Overall Sensitivity: 97,5% (95% CI 95,5% – 99,4%)

Overall Specificity: 99,7% (95% CI 99,4% – 100%)

Sensitivity

95% lower limit






Specificity

→ 95% lower limit

4th & 5th PT - Summary



- 20 labs participated and fully performed to both PT 4th & 5th
- 8 labs participated to both the PTs but reported errors in one of the two PTs
- 1 lab fully performed but partecipated only to PT 5th
- 3 labs had unsatisfactory performances in both PT 4th & 5°
- 3 labs didn't partecipate in both PTs

4th & 5th PT – Concluding remarks

Strenghts

- Web-based reporting
- Further increase in the participation and improvement in performances
- Acquisition of the International Standard for VTEC detection in food
- Most NRLs can identify VTEC genotypes
- Most NRLs can identify the "top-five" VTEC serogroups

Weakness

- 3 NRLs did not participate
- 1 NRL did not perform serotyping even for O157
- cross-contamination of samples for the 5th

Lessons learnt from the 4th and 5th PTs

Corrective actions on the EU-RL VTEC side

- Set the temperature reporting as mandatory in the web portal.
- Prepare vials for each strain in different days/spaces to avoid cross-contaminations.
- Increase the number of tests for homogeneity assessment in order to reveal problems.

Under-proficiency management

- Advice and support continuously provided to NRLs by individual contacts (Phone, E-mail)
- The training program for the NRLs will continue.
- A standard 1-week training program has been developed.
- Ad hoc visits will be done to the Labs with recurrent under-proficiency

One week-training program at EU RL-VTEC

VTEC identification and characterisation by molecular methods-Hands-on Approach-

Day 1

- introduction to conventional PCR.
- Inocula of test strains and control strains
- Overview on the activities of the EU RL-VTEC

Day 2

- Template preparation and introduction to Real Time PCR
- Conventional PCR for STEC detection and typing (CRL_Method 01 28_04_08)
- Molecular serogrouping by conventional PCR

Day 3

- Agarose gel electrophoresis
- discussion on the conventional PCR results
- Test and control strains DNA preparation for Real Time PCR

Day 4

 Real Time PCR according to the ISO TS for the identification and characterisation of using different Real Time technologies.


Day 5

Evaluation of the results obtained and general discussion on the activities carried out

EU RL main objectives in the 1st mandate:


- Establish and consolidate the network of NRLs
- Spread knowledge and technical skill on VTEC detection, identification and typing, particularly non-O157 VTEC
- Increase the coverage in the EU with respect to VTEC detection in food
- Build up an harmonized system for monitoring of VTEC in food and animal populations (zoonoses directive).


Serogroup Determination



Proficiency tests

Molecular Methods









Thank you!

For your attention

And for the efforts devoted to the success of the VTEC Network!